



# Specification

Model	Efficiency, %	Energy resolution		Peak/Compton ratio	Peak Shape	
		122 keV, (eV)	1.33 MeV, (keV)		FW.1M FWHM	FW.02M FWHM
GCD - 10 175	10	825	1.75	41:1	1.9	2.65
GCD - 15 180	15	825	1.80	46:1	1.9	2.65
GCD - 20 180	20	850	1.80	51:1	1.9	2.65
GCD - 25 185	25	850	1.85	55:1	1.9	2.65
GCD - 30 185	30	875	1.85	58:1	1.9	2.65
GCD - 35 190	35	875	1.90	60:1	1.9	2.65
GCD - 40 190	40	895	1.90	62:1	1.9	2.65
GCD - 50 190	50	895	1.90	64:1	1.9	2.65
GCD - 60 200	60	1000	2.00	68:1	2.0	3.00
GCD - 70 200	70	1000	2.00	73:1	2.0	3.00
GCD - 80 210	80	1000	2.10	77:1	2.0	3.00
GCD - 100 220	100	1000	2.10	81:1	2.0	3.00
GCD - 120 220	120	1000	2.10	83:1	2.0	3.00
GCD - 140 220	140	1100	2.20	86:1	2.0	3.00
GCD - 160 220	160*	1150	2.20	88:1	2.0	3.00

\* Detectors with higher efficiency are available



## P-type HPGe Coaxial Detectors GCD (Liquid Nitrogen cooled)

### Application

Detection of Gamma-rays in nuclear energetics and environmental control, in industry and scientific research, in medicine and other applications.

### Complete set

- HPGe coaxial detector
- Preamplifier with cooled input stage
- Different cryostat modifications are available

### Accessories

- Digital or Analog-Digital Multichannel Analyzer
- Analytical software for quantitative and qualitative analysis
- Liquid nitrogen storage and filling system

### Features

- 10% - 160% efficient HPGe coaxial detectors are available
- Energy range from 40 keV to 10 MeV with P-type HP(Ge) Detector
- Well type detectors are available
- High energy rate up to 200000 MeV/s
- Excellent peak symmetry
- Aluminum, Beryllium or Carbon-fiber input window
- Detection of radiation in any spatial orientation depending on cryostat modification
- Manufacture in variable cryostat design is possible
- HV supply protection if detector is warm
- High count rate indicator

Baltic Scientific Instruments  
Ganību Dambis 26  
P.O. box 33, Rīga  
LV - 1005, Latvia

Phone: (+371) 67383947  
Fax: (+371) 67382620  
Email: sales@bsi.lv  
www.bsi.lv

## Plenty of cryostat geometries available

